AAF do BRASIL Produtos Odontológicos Ltda.
Technician Responsible: Fabiane B. L. Santos - CRF/PR 022320
ANVISA registration: 80432410071

INSTRUCTIONS FOR USE

MOUTH OPENER

1. PRODUCT DESCRIPTION

Product designed to assist in opening the patient's mouth for dental procedures.

2. PRESENTATION

Mouth Opener S (Child), M (Adolescent), and L (Adult), packed with 1, 2, 10, 50, or 100 units.

3. COMPOSITION

Silicone.

4. INDICATION FOR USE

It is indicated for dental use to keep the patient's mouth open, facilitating the dental procedure.

5. CONSERVATION/STORAGE

Keep the product in a dry, clean, airy place at room temperature.

6. INSTRUCTIONS FOR USE

- Select the Mouth Opener with the most appropriate size for your patient;
- Place it between the teeth, so that the patient bites the serrated side and keeps the mouth open for dental procedures;
- After use, the opener must be cleaned with enzymatic soap in an ultrasonic bath and then sterilized.

7. CONTRAINDICATIONS

Should not be used in patients with sensitivity to silicone.

8. WARNINGS

The Mouth Opener is not sterile and must be sterilized before each use. Sterilization must be done in an autoclave at 121°C for 30 minutes. The product has a validity of 50 sterilization cycles. The product must be discarded according to the current legislation of the country.

9. VALIDITY

The AF do Brasil Mouth Opener is valid for 5 years from the manufacturing date.